

**CB TEST CERTIFICATE CERTIFICAT D'ESSAI OC**Product  
Produit

Medical Power Supply (for building-in)

Name and address of the applicant  
Nom et adresse du demandeurEOS Power India PVT Ltd.  
Unit #57, SDF II, Seepz-Sez, Andheri (East)  
Mumbai, 400 096  
IndiaName and address of the manufacturer  
Nom et adresse du fabricantEOS Power India PVT Ltd.  
Unit #57, SDF II, Seepz-Sez, Andheri (East)  
Mumbai, 400 096  
IndiaName and address of the factory  
Nom et adresse de l'usineEOS Power India PVT Ltd.  
Unit #57, SDF II, Seepz-Sez, Andheri (East)  
Mumbai, 400 096  
IndiaNote: When more than one factory, please report on page 2  
Note: Lorsque il y a plus d'une usine, veuillez utiliser la deuxième page Additional information on page 2Ratings and principal characteristics  
Valeurs nominales et caractéristiques principalesInput: 1.7 A Max, 100-240 Vac, 47-63 Hz  
Outputs: (See General Product Description)  
Class IITrademark (if any)  
Marque de fabrique (si elle existe)Type of Manufacturer's Testing Laboratories used  
Type de programme du laboratoire d'essais constructeur

CTF Stage 3

Model / Type Ref.  
Ref. De type

eZVCXXYY ZZt WW V M

Additional information (if necessary may also be reported on page 2)

 Additional information on page 2

Les informations complémentaires (si nécessaire, peuvent être indiqués sur la deuxième page)

A sample of the product was tested and found to be in conformity with

Un échantillon de ce produit a été essayé et a été considéré conforme à la

IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012

As shown in the Test Report Ref. No. which forms part of this Certificate

437799

Comme indiqué dans le Rapport de tests numéro de référence qui constitue partie de ce Certificat

This CB Test Certificate is issued by the National Certification Body  
Ce Certificat de test OC est établi par l'Organisme National de CertificationPhilip Pedersen vei 11,  
NO-1366 Lysaker, Norway

Date: 20-07-2021

Signature: Jiyea Gim  
Certification Department

**Additional information(if necessary)****Information complémentaire (si nécessaire)**

XX= denotes the rated output wattage designation, which can be '20' to '65', with an interval of 1. YY = Blank or alphabetical Sub Code (e.g. 30/45 -FS, 55/65 - SG/NS) ZZ= denote the standard rated output voltage designation which can be '12' to '48' or '12.0' to '48.0' in 0.1V increments t= blank or alpha or numeric character denoting minor circuit or construction modifications that do not compromise the overall safety of the product WW=Cord Type V (I/P Option) =8 – C8(Class II) M= Denotes Medical



Philip Pedersen vei 11,  
NO-1366 Lysaker, Norway

Date: 20-07-2021

Jiyea Gim

Signature: Jiyea Gim  
Certification Department