

Product Medical Power Supply

Applicant EOS Power India PVT Ltd.
Unit #57, SDF II, Seepz-Sez, Andheri (East)
Mumbai 400 096
India

Manufacturer EOS Power India PVT Ltd.
Unit #57, SDF II, Seepz-Sez, Andheri (East)
Mumbai 400 096
India

Factory EOS Power India PVT Ltd.
Unit #57, SDF II, Seepz-Sez, Andheri (East)
Mumbai 400 096
India

Ratings 1.7 A Max 100-240Vac 47-63Hz

Trade mark



Model / Type Ref. eZVCXXYY ZZt WW V M

Principal characteristics Cl. I. More information see page 2

A sample of the product was tested and found to be in conformity with MED EN 60601-1:2006;A1

Validity This certificate documents conformity with the standards shown, and also applies as license for use of Nemkos name and certification mark. The certificate and license is valid as long as the applicable conditions are complied with, and provided that any changes to the product are notified to Nemko for acceptance prior to implementation.
New standards or amendments to the standards may imply that the product design must be updated and/or that re-testing and re-certification is necessary.

Additional information See page 2

Compliance with requirements regarding building-in, protection against electric shock and Electromagnetic Compatibility (EMC) must be checked when the equipment is built-in a completed product or forms a part of a complete system.

Additional model(s) No additional models

Date of issue 20-07-2021



Juan Z. Saussey
Certification Department

Nemko AS

Philip Pedersens vei 11, 1366 Lysaker, Norway
TEL +47 22 96 03 30 EMAIL info@nemko.com
ENTERPRISE NUMBER NO974404532

Additional information

Output see the test report. In model name, XX= denotes the rated output wattage designation, which can be '20' to '65', with interval of 1. YY = Blank or alphabetical Sub Code (e.g. 30/45 -FS, 55/65 - SG/NS). ZZ= denote the standard rated output voltage designation which can be '12' to '48' or '12.0' to '48.0' in 0.1V increments t= blank or alpha or numeric character denoting minor circuit or construction modifications that do not compromise the overall safety of the product WW=Cord Type V (I/P Option) =6 - C6 (Class I), 4 - C14 (Cl. I); M= Denotes Medical

This certificate replaces the certificate issued 07-07-2021, due to the correction of the misprints

Date of issue 20-07-2021



Juan Z. Saussey

Certification Department

Nemko AS

Philip Pedersens vei 11, 1366 Lysaker, Norway
TEL +47 22 96 03 30 EMAIL info@nemko.com
ENTERPRISE NUMBER NO974404532