



STANDARD  
MODIFIED  
CUSTOMIZED  
POWER SOLUTIONS

## DECLARATION OF CONFORMITY



### The Company

**Name:** EOS Power India Private Ltd  
**Address:** Unit no #57, SDF II, Seepz, Andheri (East), Mumbai 400 096, India

declares under its sole responsibility that the product(s) to which this declaration relates is in conformity with the following directives, standard(s) and other normative document(s).

**Product:** Switch Mode Power Supply for building-in

**Models:** eZVCXXYY ZZt WW V M

XX= denotes the rated output wattage designation, which can be “20” to “65” , with interval of 1.

YY = Blank or alphabetical Sub Code (e.g. 30/45 -FS, 55/65 - SG/NS)

ZZ= denote the standard rated output voltage designation which can be “12” to “48” or “12.0” to “48.0” in 0.1V increments

t= blank or alpha or numeric character denoting minor circuit or construction modifications that do not compromise the overall safety of the product

WW=Cord Type

V (I/P Option) =6 - C6 (Class I), 4 - C14 (Class I)

M= Denotes Medical

**Quality:** ISO 9001:2015

**Medical Device Directive:** 93/42/EEC

**Product Safety Standard(s):** **IEC 60601-1:2005/A1:2012**  
**EN 60601-1 Edition 3.0 + AM1**  
(Licensed by a Notified Body to the European Union)

**EMC Directive:** 2014/30/EU Electromagnetic Compatibility

**Standard:** The Product has been tested as per standard **EN 60601-1-2 :2015\***

**RoHS Directive:** DIR 2011/65/EU and amendment (EU) 2015/863 (RoHS Directive) on the restriction of the use of certain hazardous substances in electrical & electronic equipment (EEE)

**Signed:**

**Date of Issue:** February 2022

**Name:** Vilas Redkar

**Title of Authority:** AGM -Quality

**Supplementary Information:**

The technical documentation is maintained by EOS Power India Pvt. Ltd., the signatory of this document.

\*Note: This product is classified as component type for building-in use. These products were tested as part of a standard system configuration with acceptable results. The EMC characteristic and safety approval of the final application has to be performed by the end-product manufacturer in accordance to the applicable standards for the end-product