

We, EOS Power India Private Ltd, Unit no #57, SDF II, Seepz, Andheri (East), Mumbai 400 096, India, a manufacturer declare under our sole responsibility that the products

Power Supply Model(s): MULP180-ZVXX-YYY

Z= 1 is for models without PG/PF and 0 is with PG/PF.

V= 0 to 9 denoting the output connector.

X = any numeric from 0-9 (Representing the output voltage which can be any voltage in the range 12V to 58V, max current as stated in ref table)

Y = any alpha or numeric character or blank (denotes minor output variation and /or minor SELV circuit variations)

to which this declaration relates, is/are in compliance with the following document(s):

Quality Standard(s): **ISO 9001**

UK Legislation: **Medical Devices Regulations 2002 (SI 2002 No 618, as amended)
(UK MDR 2002):**


Product Safety Standard(s): **IEC/EN 60601-1 Edition 3.0 + AM1**
(Licensed by a Notified Body to the European Union)

UK Legislation: **S.I. 2012 No. 3032: The Restriction of the Use of Certain
Hazardous Substances in Electrical and Electronic Equipment Regulations 2012
EN IEC 63000:2018**

UK Legislation: **Medical Devices Regulations 2002 (SI 2002 No 618, as amended)
(UK MDR 2002):
EN 60601-1-2 :2015**

This component power supply is intended exclusively for inclusion within other equipment. The power supply is considered Class I and must be connected to a reliable earth grounding system

Approved by:



(Manufacturer)
Vilas Redkar
AGM Quality

Mumbai, India
(Place)

February 14, 2022,
(Date)

UK Representative:



(Manufacturer)
Cliff Gore
European Sales Director
Bel Power Solutions

Maidstone, UK
(Place)

February 14, 2022
(Date)