



STANDARD
MODIFIED
CUSTOMIZED
POWER SOLUTIONS

UKCA DECLARATION OF CONFORMITY



We, EOS Power India Private Ltd, Unit no #57, SDF II, Seepz, Andheri (East), Mumbai 400 096, India, a manufacturer declare under our sole responsibility that the products

Power Supply Model(s): eZVCXXYY ZZt WW V M

XX = denotes the rated output wattage designation, which can be “20” to “65, with interval of 1.

YY = Blank or alphabetical Sub Code (e.g. 30/45 -FS, 55/65 - SG/NS).

ZZ = denote the standard rated output voltage designation which can be “12” to “48” or “12.0” to “48.0” in 0.1V increments

t = blank or alpha or numeric character denoting minor circuit or construction modifications that do not compromise the overall safety of the product

WW = Cord Type V (I/P Option) =6 - C6 (Class I), 4 - C14 (Class I)

M= Denotes Medical

to which this declaration relates, is/are in compliance with the following document(s):

Quality Standard(s): **ISO 9001**

UK Legislation: **Medical Devices Regulations 2002 (SI 2002 No 618, as amended)
(UK MDR 2002):**

Product Safety Standard(s): **IEC/EN 60601-1 Edition 3.0 + AM1**
(Licensed by a Notified Body to the European Union)

UK Legislation: **S.I. 2012 No. 3032: The Restriction of the Use of Certain
Hazardous Substances in Electrical and Electronic Equipment Regulations 2012
EN IEC 63000:2018**

UK Legislation: **Medical Devices Regulations 2002 (SI 2002 No 618, as amended)
(UK MDR 2002):
EN 60601-1-2 :2015**

This component power supply is intended exclusively for inclusion within other equipment. The power supply is considered Class I and must be connected to a reliable earth grounding system

Approved by:

(Manufacturer)
Vilas Redkar
AGM Quality

Mumbai, India

(Place)

February 14, 2022,

(Date)

UK Representative:

(Manufacturer)

Cliff Gore
European Sales Director
Bel Power Solutions

Maidstone, UK

(Place)

February 14, 2022

(Date)