

We, EOS Power India Private Ltd, Unit no #57, SDF II, Seepz, Andheri (East), Mumbai 400 096, India, a manufacturer declare under our sole responsibility that the products

**Power Supply Model(s): ) : eZVC XX YY ZZt WW V M**

XX = denotes the rated output wattage designation, which can be “20” to “65”,, with interval of 1.

YY = Blank or alphabetical Sub Code (e.g. 30/45 -FS, 55/65 - SG/NS).

ZZ = denote the standard rated output voltage designation which can be “12” to “48” or “12.0” to “48.0”in 0.1V increments

t = blank or alpha or numeric character denoting minor circuit or construction modifications that do not compromise the overall safety of the product

WW = Output Cord Type

V (I/P Option) = 8- C8 (Class II)

M = Denotes Medical

to which this declaration relates, is/are in compliance with the following document(s):

UK Legislation: **Medical Devices Regulations 2002 (SI 2002 No 618, as amended)  
(UK MDR 2002):**

Product Safety Standard(s): **IEC/EN 60601-1 Edition 3.0 + AM1**  
(Licensed by a Notified Body to the European Union)

UK Legislation: **S.I. 2012 No. 3032: The Restriction of the Use of Certain  
Hazardous Substances in Electrical and Electronic Equipment Regulations  
2012**  
**EN IEC 63000:2018**

UK Legislation: **Medical Devices Regulations 2002 (SI 2002 No 618, as amended)  
(UK MDR 2002):**  
**EN 60601-1-2 :2015**

This component power supply is intended exclusively for inclusion with other equipment

Approved by:



(Manufacturer)

Vilas Redkar  
AGM Quality

UK Representative:



(Manufacturer)

Cliff Gore  
European Sales Director  
Bel Power Solutions

**Mumbai, India**

(Place)

**February 14, 2022,**

(Date)

**Maidstone, UK**

(Place)

**February 14, 2022**

(Date)