

We, EOS Power India Private Ltd, Unit no #57, SDF II, Seepz, Andheri (East), Mumbai 400 096, India, a manufacturer declare under our sole responsibility that the products

Power Supply Model(s): MFLS250- ZXVV-YYYY-MMMM

Z = no of Outputs, 1 for single Output & 2 with Standby 5Vsb

X = numeric character for denoting type of connector (e.g.0= For Screw Terminal, 3= PCB header)

V = any numeric from 0-9 (Representing the output voltage which can be any voltage in the range 12V to 58V

Y = any alpha, numeric or alphanumeric character or blank (denotes minor output variation and /or minor SELV circuit variations);

M= any alpha, numeric or alphanumeric character or blank, denotes type of metal accessories (e.g. U=U channel,

L= L bracket, CK= Cover-kit)

to which this declaration relates, is/are in compliance with the following document(s):

Quality Standard(s): **ISO 9001**

UK Legislation: **Medical Devices Regulations 2002 (SI 2002 No 618, as amended)
(UK MDR 2002):**


Product Safety Standard(s): **IEC/EN 60601-1 Edition 3.0 + AM1**
(Licensed by a Notified Body to the European Union)

UK Legislation: **S.I. 2012 No. 3032: The Restriction of the Use of Certain
Hazardous Substances in Electrical and Electronic Equipment Regulations 2012
EN IEC 63000:2018**

UK Legislation: **Medical Devices Regulations 2002 (SI 2002 No 618, as amended)
(UK MDR 2002):
EN 60601-1-2 :2015**

This component power supply is intended exclusively for inclusion within other equipment. The power supply is considered Class I and must be connected to a reliable earth grounding system

Approved by:



(Manufacturer)
Vilas Redkar
AGM Quality

Mumbai, India
(Place)

February 14, 2022,
(Date)

UK Representative:



(Manufacturer)
Cliff Gore
European Sales Director
Bel Power Solutions

Maidstone, UK
(Place)

February 14, 2022
(Date)