

Additional model(s)

Product

Medical Power Supply (for building-in)

Pos. No

1

Model / Type Ref.

MVPS800-1ZXX-YYYY-R

Trade mark (if different from page 1)

Rating

Input: 8.0 A Max, 100-240 Vac, 47-63 Hz, Cl. I

Output: 800 W max (Refer to General product information in report 360643 for output rating conditions and variable definition)

Principal characteristics

Where: Z = can be 'S', 'T' or 0 (Representing type of top cover, S= Slotted, T= Plain and 0= no top cover)

X = any numeric from 0-9 (Representing the output voltage which can be any voltage in the range 12V to 58V, max current as stated in ref table)

Y = any alpha, numeric or alphanumeric character or blank (denotes minor output variation and /or minor SELV circuit variations)

R= with built-in Redundant diode (Optional)

Date of issue 29-01-2019



Okhyun Jeon

Certification Department

Nemko AS

Gaustadalléen 30, P.O. Box 73 Blindern, 0314 Oslo, Norway

TEL +47 22 96 03 30 FAX +47 22 96 05 50 EMAIL info@nemko.com

ENTERPRISE NUMBER