

CB TEST CERTIFICATE CERTIFICAT D'ESSAI OCProduct
ProduitName and address of the applicant
Nom et adresse du demandeurName and address of the manufacturer
Nom et adresse du fabricantName and address of the factory
Nom et adresse de l'usineNote: When more than one factory, please report on page 2
Note: Lorsque il y plus d'une usine, veuillez utiliser la deuxième pageRatings and principal characteristics
Valeurs nominales et caractéristiques principalesTrademark (if any)
Marque de fabrique (si elle existe)Type of Manufacturer's Testing Laboratories used
Type de programme du laboratoire d'essais constructeurModel / Type Ref.
Ref. De typeAdditional information (if necessary may also be reported on page 2)
Les informations complémentaires (si nécessaire, peuvent être indiqués sur la deuxième pageA sample of the product was tested and found to be in conformity with
Un échantillon de ce produit a été essayé et a été considéré conforme à laAs shown in the Test Report Ref. No. which forms part of this Certificate
Comme indiqué dans le Rapport desais numéro de référence qui constitue partie de ce CertificatThis CB Test Certificate is issued by the National Certification Body
Ce Certificat desai OC est établi par l'Organisme **National de Certification**

Medical power supply for (building-in)

EOS Power India PVT Ltd.
Unit #57, SDF II, Seepz-Sez
Andheri (East), Mumbai 400 096
IndiaEOS Power India PVT Ltd.
Unit #57, SDF II, Seepz-Sez
Andheri (East), Mumbai 400 096
IndiaEOS Power India PVT Ltd.
Unit #57, SDF II, Seepz-Sez
Andheri (East), Mumbai 400 096
India Additional information on page 2

AC Input: 2.0 A Max, 100- 240 V~, 47- 63 Hz, CI. I


LFMWLP75-1XXX-YYY-W, LFMWLP65-1XXX-YYY-W or
LFMWLP45-1XXX-YYY-WOutput Rating see the test report. In model name, the: "X" (first) = 0 or 3;
"XX" (second and third) = 01 to 99; "Y" = 0-9, A-Z, a-z or blank, W is blank
for MOPP or 1 for MOOP for CI. I Additional information on page 2

IEC 60601-1(ed.3);am1

294295

Gautadalléen 30
NO-0373 Oslo, Norway

Date: 27-10-2015

Signature: Juan Z. Kleppenes
Certification Department