

# CERTIFICATE OF COMPLIANCE

**Certificate Number** 20151106-E173812  
**Report Reference** E173812-A28-UL  
**Issue Date** 2015-NOVEMBER-06

**Issued to:** EOS POWER INDIA PVT LTD  
UNIT 57 SDF-II  
SEEPZ ANDHERI (E)  
MUMBAI MH 400096 INDIA

**This is to certify that  
representative samples of**

COMPONENT - POWER SUPPLIES, MEDICAL AND  
DENTAL

Medical Power Supply (For Building-in)  
LFMWLP120-1XXX-YYY-II, LFMWLP75-1XXX-YYY-II,  
LFMWLP65-1XXX-YYY-II or LFMWLP45-1XXX-YYY-II,  
(Where: "X" (first) = 0 or 3; "XX" (second and third) = 01 to  
99; "Y" = 0-9, A-Z, a-z or blank)

Have been investigated by UL in accordance with the  
Standard(s) indicated on this Certificate.

**Standard(s) for Safety:** ANSI/AAMI ES60601-1 and CAN/CSA-C22.2 No. 60601-1,  
Medical Electrical Equipment - Part 1: General  
Requirements for Basic Safety and Essential Performance

**Additional Information:** See the UL Online Certifications Directory at  
[www.ul.com/database](http://www.ul.com/database) for additional information

Only those products bearing the UL Certification Mark should be considered as being covered by UL's  
Certification and Follow-Up Service.

Recognized components are incomplete in certain constructional features or restricted in performance  
capabilities and are intended for use as components of complete equipment submitted for investigation rather  
than for direct separate installation in the field. The final acceptance of the component is dependent upon its  
installation and use in complete equipment submitted to UL LLC.

Look for the UL Certification Mark on the product.



Bruce Mahrenholz, Assistant Chief Engineer, Global Inspection and Field Services  
UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please  
contact a local UL Customer Service Representative at [www.ul.com/contactus](http://www.ul.com/contactus)

