


| | |
|---|--|
| Product | Medical Power Supply (for building-in) |
| Applicant | EOS Power India PVT Ltd. Unit #57, SDF II, Seepz-Sez, Andheri (East) Mumbai 400 096 India |
| Manufacturer | EOS Power India PVT Ltd. Unit #57, SDF II, Seepz-Sez, Andheri (East) Mumbai 400 096 India |
| Factory | EOS Power India PVT Ltd. Unit #57, SDF II, Seepz-Sez, Andheri (East) Mumbai 400 096 India |
| Ratings | 3.0A Max. 100-240VAC 47-63Hz |
| Trade mark |  |
| Model / Type Ref. | LFMWLT200-ZXXX-YYY-W2 |
| Principal characteristics | Cl. II. DC output see the test report. More information see the additional information |
| A sample of the product was tested and found to be in conformity with | MED EN 60601-1:2006;A1 |
| Validity | This certificate documents conformity with the standards shown, and also applies as license for use of Nemkos name and certification mark. The certificate and license is valid as long as the applicable conditions are complied with, and provided that any changes to the product are notified to Nemko for acceptance prior to implementation. New standards or amendments to the standards may imply that the product design must be updated and/or that re-testing and re-certification is necessary. |
| Additional information | See page 2 Compliance with requirements regarding building-in, protection against electric shock and Electromagnetic Compatibility (EMC) must be checked when the equipment is built-in a completed product or forms a part of a complete system. |
| Additional model(s) | (1) See page 2 |

Date of issue 06-01-2016



Juan Z. Kleppenes
Certification Department

Nemko AS
Gaustadalléen 30, P.O. Box 73 Blindern, 0314 Oslo, Norway
TEL +47 22 96 03 30 FAX +47 22 96 05 50 EMAIL info@nemko.com
ENTERPRISE NUMBER NO974404532