

UL TEST REPORT AND PROCEDURE

Standard:	UL 60601-1, 1st Edition, 2006-04-26 (Medical Electrical Equipment, Part 1: General Requirements for Safety) CAN/CSA-C22.2 No. 601.1-M90, 2005 (Medical Electrical Equipment - Part 1: General Requirements for Safety)
Certification Type:	Power Supplies, Medical and Dental
CCN:	QQHM2, QQHM8
Product:	Power supply
Model:	Model series LFMWLT200-ZXXX-YYY-W, where Z can be any number between 1 to 4, XXX can be any number from 001 to 999, YYY can be any alphanumeric number or blank, W can be either blank for Class I or 2 for Class II.
Rating:	<p>LFMWLT200-1000: Input: 100-240 V~, 3.0 Amax, 47-63 Hz Output: + 5 Vdc, 35 A; +12 Vdc, 0.5 A (fan); + 5 V, 1 A (standby), 175 Wmax</p> <p>LFMWLT200-1001: Input: 100-240 V~, 3.0 Amax, 47-63 Hz Output: + 12 Vdc, 16.67 A; +12 Vdc, 0.5 A (fan); + 5 V, 1 A (standby), 200 Wmax</p> <p>LFMWLT200-1002: Input: 100-240 V~, 3.0 Amax, 47-63 Hz Output: + 15 Vdc, 13.33 A; +12 Vdc, 0.5 A (fan); + 5 V, 1 A (standby), 200 Wmax</p> <p>LFMWLT200-1003: Input: 100-240 V~, 3.0 Amax, 47-63 Hz Output: + 24 Vdc, 8.33 A; +12 Vdc, 0.5 A (fan); + 5 V, 1 A (standby), 200 Wmax</p> <p>LFMWLT200-1004: Input: 100-240 V~, 3.0 Amax, 47-63 Hz Output: + 48 Vdc, 4.17 A; +12 Vdc, 0.5 A (fan); + 5 V, 1 A (standby), 200 Wmax</p>
Applicant Name and Address:	RED ROCKET INC UNIT 200 1540 CORNERSTONE BLVD DAYTONA BEACH FL 32117 UNITED STATES


This is to certify that representative samples of the products covered by this Test Report have been investigated in accordance with the above referenced Standards. The products have been found to comply with the requirements covering the category and the products are judged to be eligible for Follow-Up Service under the indicated Test Procedure. The manufacturer is authorized to use the UL Mark on such products which comply with this Test Report and any other applicable requirements of Underwriters Laboratories Inc. ('UL') in accordance with the Follow-Up Service Agreement. Only those products which properly bear the UL Mark are considered as being covered by UL's Follow-Up Service under the indicated Test Procedure.

The applicant is authorized to reproduce the referenced Test Report provided it is reproduced in its entirety.

UL authorizes the applicant to reproduce the latest pages of the referenced Test Report consisting of the first page of the Specific Technical Criteria through to the end of the Conditions of Acceptability.

Any information and documentation involving UL Mark services are provided on behalf of Underwriters Laboratories Inc. (UL) or any authorized licensee of UL.

Prepared by: Ahmed El-Gawish
Underwriters Laboratories Inc.
Reviewed by: Richard Dolle
Underwriters Laboratories Inc.

Handwritten signatures of Ahmed El-Gawish and Richard Dolle. The signature of Ahmed El-Gawish is written in black ink and is positioned above the signature of Richard Dolle. The signature of Richard Dolle is also written in black ink and is positioned below the signature of Ahmed El-Gawish.