


Product	Medical Power Supply (for building-in)
Applicant	EOS Power India PVT Ltd. Unit #57, SDF II, Seepz-Sez, Andheri (East) Mumbai 400 096 India
Manufacturer	EOS Power India PVT Ltd. Unit #57, SDF II, Seepz-Sez, Andheri (East) Mumbai 400 096 India
Factory	EOS Power India PVT Ltd. Unit #57, SDF II, Seepz-Sez, Andheri (East) Mumbai 400 096 India
Ratings	1.5A 100-240V AC 47-63Hz
Trade mark	
Model / Type Ref.	LFMWLT60-ZXXX-YYY-WII
Principal characteristics	Cl. II. DC outputs: Refer to test report. The "Z" in model designation may be 1, 2 or 3, representing number of outputs. The "X" may be any numeric character, denoting configuration. The "Y" may be any alphanumeric character or blank, denoting minor output variation and/or minor SELV circuit variation. "W" may be 1 or blank and "II" denote class II construction.
A sample of the product was tested and found to be in conformity with	MED EN 60601-1:2006;A1
Validity	This certificate documents conformity with the standards shown, and also applies as license for use of Nemkos name and certification mark. The certificate and license is valid as long as the applicable conditions are complied with, and provided that any changes to the product are notified to Nemko for acceptance prior to implementation. New standards or amendments to the standards may imply that the product design must be updated and/or that re-testing and re-certification is necessary.
Additional information	Compliance with requirements regarding building-in, protection against electric shock and Electromagnetic Compatibility (EMC) must be checked when the equipment is built-in a completed product or forms a part of a complete system.
Additional model(s)	(3) See page 2

Date of issue 12-01-2016



Skule Moe
Certification Department

Nemko AS
Gaustadalléen 30, P.O. Box 73 Blindern, 0314 Oslo, Norway
TEL +47 22 96 03 30 FAX +47 22 96 05 50 EMAIL info@nemko.com
ENTERPRISE NUMBER NO974404532



Order No. 294630

CERTIFICATE

No. P16220664



Page 2 of 2

Additional model(s)

Product	Medical Power Supply (for building-in)
Pos. No	1
Model / Type Ref.	LFMWLT40-ZXXX-YYY-WII
Trade mark (if different from page 1)	
Rating	1.5A 100-240V AC 47-63Hz
Principal characteristics	Cl. II. DC outputs: Refer to test report. The "Z" in model designation may be 1, 2 or 3, representing number of outputs. The "X" may be any numeric character, denoting configuration. The "Y" may be any alphanumeric character or blank, denoting minor output variation and/or minor SELV circuit variation. "W" may be 1 or blank and "II" denote class II construction.